



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/506,942	04/05/2005	Christoph Von Mandach	9593.18619-PCT US	6984
26308	7590	06/14/2006	EXAMINER	
RYAN KROMHOLZ & MANION, S.C. POST OFFICE BOX 26618 MILWAUKEE, WI 53226			KILKENNY, PATRICK J	
			ART UNIT	PAPER NUMBER
			3732	

DATE MAILED: 06/14/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

10/506,942

Applicant(s)

VON MANDACH, CHRISTOPH

Examiner

Patrick J. Kilkeny

Art Unit

3732

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 30 March 2006.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 24-49 is/are pending in the application.
- 4a) Of the above claim(s) 28-49 is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 24-27 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- ☐ Notice of References Cited (PTO-892)
- ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____
- ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____
- ☐ Notice of Informal Patent Application (PTO-152)
- ☐ Other: _____

DETAILED ACTION

Election/Restrictions

Claims 28-49 are withdrawn from further consideration pursuant to 37 CFR 1.142(b) as being drawn to a nonelected apparatuses for use in the application of an orthodontic fastening part, there being no allowable generic or linking claim. Election was made **without** traverse in the reply filed on 3/27/2006.

Applicant's election without traverse of Invention I, claims 24-27, drawn to a process for the application of and orthodontic fastening part in the reply filed on 3/27/2006 is acknowledged.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(a) the invention was known or used by others in this country, or patented or described in a printed publication in this or a foreign country, before the invention thereof by the applicant for a patent.

Claims 24 and 25 are rejected under 35 U.S.C. 102(B) as being anticipated by Dellinger (4,360,341). Dellinger discloses a method for the application of an orthodontic fastening part (Fig.5, #18), which includes covering the fastening part with a complementary protector (Fig. 5, # 24), accommodating the fastening part/protector combination with a complementary applicator (Fig. 5, #32), depositing adhesive onto the fastening part and allowing the adhesive to cure, after which the protector and applicator are removed (Column 3, lines 58-62). The fastening part is aligned on the

tooth by the contoured ends formed on the outer surface of the applicator, which have been preformed to each tooth surface and help align the fastening part (Fig. 3, #'s 34 and 36).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claim 26 is rejected under 35 U.S.C. 103(a) as being unpatentable over Dellinger. Dellinger discloses the claimed method except for the orthodontic fastening part and the protector being attached in a preassembled manner. It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify the invention of Dellinger by integrally combining the orthodontic fastening part (Fig. 5, #18) and the protector (Fig. 5, #24), since it has been held that forming in one piece an article which has formerly been formed in two pieces and put together involves only routine skill in the art. *Howard v. Detroit Stove Works*, 150 U.S. 164 (1893).

Claim 27 is rejected under 35 U.S.C. 103(a) as being unpatentable over Dellinger in view of Podszun et al. (5,921,779). Dellinger discloses the claimed method with the exception of using a polymerization lamp for curing the dental adhesive. Podszun et al. discloses using a polymerization lamp for curing dental adhesive (Column 4, lines 61-63). Therefore, it would have been obvious for one of ordinary skill in the art at the time

Art Unit: 3732

the invention was made to utilize a polymerization lamp to cure the dental adhesive, as taught by Podszun, and as is well known in the art, in order to rapidly cure the dental adhesive and obtain a strong bond between the orthodontic fixing element and the tooth.

Response to Arguments

Applicant's arguments filed 3/30/2006 have been fully considered but they are not persuasive. Although the bracket of Dillinger is pre-aligned in the same position it was on the replica tooth, the contoured ends of the applicator (34 and 36) are still "aids" for aligning the bracket because they are portions on the outer surface of the applicator that are fit into position on the tooth. In response to applicant's argument that the references fail to show certain features of applicant's invention, it is noted that the features upon which applicant relies (i.e., a y-direction pointer, an angulation pointer, or a torque indicator) are not recited in the rejected claim(s). Although the claims are interpreted in light of the specification, limitations from the specification are not read into the claims. See *In re Van Geuns*, 988 F.2d 1181, 26 USPQ2d 1057 (Fed. Cir. 1993).

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP

§ 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.


Any inquiry concerning this communication or earlier communications from the examiner should be directed to Patrick J. Kilkenny whose telephone number is (571) 272-8684. The examiner can normally be reached on Mon-Fri, 8-5:30.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Patricia Bianco can be reached on (571) 272-4940. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3732

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.


PJK


Patrick J. Kilkenney
Art Unit 3732
June 10, 2006
6/10/06


PATRICIA BIANCO
PRIMARY EXAMINER
6/12/06